

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 1/22/2020 Instrument Serial Number: 100422

Instrument Location: TBER

Supervisor Name and Agency: Colleen O'Bryant / SCDL

Describe the Status Change or Issue with the Instrument:

In correct external standard expiration date listed on VOCs from 12/5/2019 - 1/3/2020.
See email for specifics. Lot AB826101 exp 9/18/2020
Scott Gardner corrected the expiration date on 1/22/2020 and performed a VOC.
Dates 12/5/2019 - 1/22/2020. CAB 1/23/2020

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: _____

For Use by SCDL	Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

O'Bryant, Colleen S (DPS)

From: O'Bryant, Colleen S (DPS)
Sent: Tuesday, January 21, 2020 1:44 PM
To: scott.gardner.14@us.af.mil
Cc: brandi.barnett@alaska.gov; Walton, Derek J (DPS)
Subject: 100422

Scott,

DataMaster DMT serial number 100422 has the wrong expiration date currently programed for the external standard tank.

On 12/05/2019 you did a tank change and used lot number: AG826101 with expiration date: 12/18/2020. The correct expiration date is 9/18/2020. Since 12/05/2019 three VOCs have been completed with the incorrect expiration date as well as a handful of subject tests.

As soon as possible please run a VOC on 100422 and correct the expiration date. I will hand correct the three VOCs with incorrect expiration dates.

Colleen O'Bryant

**Forensic Scientist III
Blood and Breath Alcohol Program
Alaska Scientific Crime Detection Laboratory
4805 Dr. Martin Luther King Jr. Ave
Anchorage, Alaska 99507**

**907-269-5783 (direct)
907-350-2111 (cell)**