

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 8/29/17 Instrument Serial Number: 100415

Instrument Location: Sitka PD

Supervisor Name and Agency: Charles Foster, SCDL

Describe the Status Change or Issue with the Instrument:


If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: _____

For Use by SCDL

Additional Notes

During the 8/22-8/24/2017 Breath Test Supervisor Certification class, Gary Cranford (Sitka PD) informed me that the breath hose from their previous DMT (100690) was installed on 100415 when it was put into service on 6/1/2017. The breath alcohol program staff has concluded a review of all breath tests administered on instrument 100415, with the breath hose of 100690 installed, and determined that all of the tests are scientifically valid and accurate. Because the breath hose contains the antenna that detects radio frequency interference, I contacted Gary Cranford on 8/29/17 to walk him through setting the RFI sensitivity and conducting an RFI detected test by keying a radio during a subject test sequence (test print out attached). Both of these activities were completed successfully. Afterwards, a Verification of Calibration testing sequence was initiated. - Charles Foster 8/29/17 

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100415

Date: 08/29/2017
Time: 13:07:39

SUBJECT NAME:
TEST TEST
LICENSE #: RFI TEST

OPERATOR NAME:
GARY W CRANFORD
OPERATOR'S #: 8892

DEPT/AGENCY:
SITKA PD
CASE/REPORT: RFI TEST
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.59: 0.079
LOT #: AG634001
EXPIRATION: 12/05/2018

DIAGNOSTIC CHECK
BLANK TEST PASSED 13:08
INTERNAL STANDARD VERIFIED 13:08
EXTERNAL STANDARD 0.078 13:09
BLANK TEST 0.000 13:10
RFI DETECTED

W. J. R. S. I

