

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 11/5/2022

Instrument #: 100373

Instrument Location: Anchorage Jail

Supervisor Name and Agency: Jason Cusack, Anchorage Police Department

Describe the Status Change or Issue with the Instrument:

Instrument was taken out of service due to not registering when a breath sample was being provided. I attempted to restart the datamaster as well as re-install the breath hose without any success in getting it to work again.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: [Click or tap here to enter text.](#)

Out of Service Date/Time: 11/5/2022 0040 hours

-----  
**For Use by SCDL**

Additional Notes

[Click or tap here to enter text.](#)

**Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)**  
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor