

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Effective: 11/10/2021

Version: 2.0

Date: 5/6/2022 _____ Instrument #: 100378 _____

Instrument Location: Palmer PD _____

Supervisor Name and Agency: Carlie Glaister - SCDL _____

Describe the Status Change or Issue with the Instrument:

When reviewing status messages during IFIR it was noticed that Palmer PD has a higher percentage of RFI status messages than average. An email was sent to the Breath Test Supervisor, Lucas Szipszky, to remind Breath Test Operators and other officers not to key radios within the vicinity of the DataMaster when it is in use.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA _____

Out of Service Date/Time: NA _____

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740