

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100699

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 01/03/2026

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 29.56: 0.079

LOT #: AG528101

EXPIRATION: 10/08/2027

TANK PRESSURE: 1141 psi

BLANK TEST 0.000 12:02

INTERNAL STANDARD ERROR 12:02

Internal Standard Error

Diagnostic Check

VERSIONS

DMT: 3.02

PIC: 3.03

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 49.0°C PASSED

Breath Tube = 44.8°C PASSED

PUMP INFO

Flow Rate = 4.158 L/M PASSED

DETECTOR INFO

PUMP ON PASSED

PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED

Filter 2 PASSED

Filter 3 PASSED

INTERNAL STANDARD PASSED

File only
01/03/26