

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 1/24/15 Instrument Serial Number: 100411
Instrument Location: Wildwood Correctional
Supervisor Name and Agency: Todd Hamilton / Kenai PD

Describe the Status Change or Issue with the Instrument:

Higher number of Interference Detected on the external standard than is normal. Also have received some internal standard status messages.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 1/24/15 @ 1430

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740