

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 4-28-2015 Instrument Serial Number: 100345
Instrument Location: King Cove PD
Supervisor Name and Agency: Christopher m Babcock

Describe the Status Change or Issue with the Instrument:

Datamaster will not pass Diagnostic Check
Detector bad

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 4-28-2015 1400

For Use by SCDL

Additional Notes

Detector overflow status message during 4/11/15 scheduled
VOE
- CBF 5/1/15

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740