

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 01/23/2024

Instrument #: 100682

Instrument Location: King Cove PD

Supervisor Name and Agency: Carlie Glaister SCDL

Describe the Status Change or Issue with the Instrument:

While technically reviewing VOCs, I noticed that the expiration # for the external standard was incorrect (listed as 5/19/2025) for 100682 VOC that occurred on 01/03/2024. I handcorrected the VOC and notified the King Cove PD supervisor to perform a tank change ASAP to correct the expiration #. No additional VOCs or breath records are impacted at this time.

Correct Tank lot # & Expiration date: AG310901, 4/19/2025 - added 3/4/2024 CKG

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: [Click or tap here to enter text.](#)

Out of Service Date/Time: [Click or tap here to enter text.](#)

For Use by SCDL

Additional Notes

[Click or tap here to enter text.](#)

Email completed form to dps.scdl.toxicology@alaska.gov.

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

Glaister, Carlie K (DPS)

From: Glaister, Carlie K (DPS)
Sent: Tuesday, January 23, 2024 2:03 PM
To: King Cove Department of Public Safety
Cc: Walton, Derek J (DPS)
Subject: Perform tank change ASAP

Importance: High

Hello, I am reviewing VOCs and the VOC performed on 1/3/2024 has the wrong expiration date associated with the external tank. So far no other VOCs or breath records are impacted. Please remove the tank and perform a tank change as soon as possible to correct the expiration date before running anything else on the DataMaster. Let me know when this is completed.

Carlie Glaister