

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100367

Received From: Seward PD

Date: 12/20/23

Analyst: C. OBryant

Reason for instrument return:

Internal Standard error

Change instrument and case location in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax

Upload

Update agency information and polling status in DMHost Calling Program

Check for VOCs

From Intoximeters

Intoximeters Repair Paperwork (attached)