

CALIBRATION FACTORS

-----  
Alaska Department of Public Safety  
DATAMASTER dmt: 100376  
-----

Date: 03/24/2020  
Time: 07:25:22

OPERATOR NAME:  
INTOX I METERS  
CERTIFICATE #: 1

Ca = 0.2000  
CAL = 0.972167    0.800 <= CAL < 1.200  
b1 = 0.0005    0.0000 <= b1 < 0.0040  
b2 = 0.0065    0.0010 <= b2 < 0.0100  
b3 = 0.0002    0.0000 <= b3 < 0.0040  
Xq = 0.0743    0.0500 <= Xq < 0.2000  
a21 = 1.172072    1.050 <= a21 < 1.300  
a31 = 0.369043    0.300 <= a31 < 0.800