

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Issued: 7/10/2015  
Effective: 7/10/2015

Version: DMCIF 2015 R1  
Status: Active

Instrument S/N 100421

Received From JSBER

Date 4/1/16

Analyst CWF

1. Reason for instrument return:

Blank & Ambient Fail errors, Modem connectivity issues

2. Change instrument and case location in JusticeTrax Breath Alcohol Database  No case

3. Ensure change in instrument status form is in JusticeTrax

4. Upload  Took two tries

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs

From NPAS

7. NPAS Repair Paperwork (attached)