

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100418

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 03/11/2017

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 30.19: 0.081
LOT #: AG608504
EXPIRATION: 03/25/2018
TANK PRESSURE: 1026 psi

| | | |
|-------------------|----------|-------|
| BLANK TEST | 0.000 | 09:25 |
| INTERNAL STANDARD | VERIFIED | 09:25 |
| EXTERNAL STANDARD | 0.085 | 09:25 |
| BLANK TEST | 0.000 | 09:26 |
| EXTERNAL STANDARD | 0.086 | 09:26 |
| BLANK TEST | 0.000 | 09:27 |
| EXTERNAL STANDARD | 0.086 | 09:27 |
| BLANK TEST | 0.000 | 09:28 |
| EXTERNAL STANDARD | 0.085 | 09:28 |
| BLANK TEST | 0.000 | 09:29 |
| EXTERNAL STANDARD | 0.086 | 09:29 |
| BLANK TEST | 0.000 | 09:30 |

Average = 0.0856
Std Dev = 0.0005

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C PASSED
Breath Tube = 46.4°C PASSED

PUMP INFO

Flow Rate = 4.587 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED

File only: Instrument being removed from service.

BMB 6/29/17