

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 10/26/16 Instrument Serial Number: 100688

Instrument Location: Hooper Bay PD

Supervisor Name and Agency: Colleen O'Bryant, SCDL

Describe the Status Change or Issue with the Instrument:

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 10/25/16 4P

For Use by SCDL

Additional Notes

Removed from service remotely via DM Host
for detector overflow status messages following
an attempted VOC. Detector voltage is elevated.
No out of service VOC.

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

System Time

2016/10/25 15:56

Temperatures (°C)

Sample Cell: 48.8 Sim. Hose 0.0
Breath Tube: 47.7 Sim. Temp 0.00

Settings (V)

Lamp: 1.75 + Save
Bias: 80.4 + Save
Cooler: 1.88 + Save
Chopper: 560 + Save

On/Off Barometer (mbar) 993

Chopper

Pump Set

Sim. Valve Tank Psr (psi) 883.62

Gas Valve

Stepper

Filter 1 Filter 1 + Quartz

Filter 2 Filter 2 + Quartz

Filter 3 Filter 3 + Quartz

Voltages (V) Flow: 1.01 Detector: 1.373

Remote VOC

Alaska Department of Public Safety
DATAMASTER dmt: 190688

Date: 10/25/2016
Time: 15:57:05

VERSIONS
DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
Breath Tube = 47.7°C PASSED

PUMP INFO
Flow Rate = 4.168 L/M PASSED

DETECTOR INFO
PUMP ON PASSED
PUMP OFF PASSED

Detector: OverFlow

Print

Cancel

Hopper Bay PD - 100688
COB 10/24/16

<input type="checkbox"/>	100689	Katzebue PD
<input type="checkbox"/>	100690	Sitka PD
<input type="checkbox"/>	100691	Valdez PD
<input type="checkbox"/>	100692	AST Aniak
<input type="checkbox"/>	100693	State Game Lab
<input type="checkbox"/>	100694	Petersburg PD