

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100365

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 03/21/2019

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.05: 0.078
LOT #: AG735001
EXPIRATION: 12/16/2019
TANK PRESSURE: 1100 psi

BLANK TEST	0.000	15:43
INTERNAL STANDARD	VERIFIED	15:43
EXTERNAL STANDARD	0.076	15:43
BLANK TEST	0.000	15:44
EXTERNAL STANDARD	0.076	15:45
BLANK TEST	0.000	15:45
EXTERNAL STANDARD	0.076	15:46
BLANK TEST	0.000	15:47
EXTERNAL STANDARD	0.076	15:47
BLANK TEST	0.000	15:48
EXTERNAL STANDARD	0.076	15:48
BLANK TEST	0.000	15:49

Average = 0.0760
Std Dev = 0.0000

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
Breath Tube = 46.2°C PASSED

PUMP INFO

Flow Rate = 4.308 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Charles R. Foster, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Charles R. Foster
Scientific Director
State Breath Alcohol Program

5/31/19

Subscribed and sworn before me this 31 day of MAY, 20 19


Notary Public
My Commission Expires With Office



DM 05/08/19