

# Alaska Scientific Crime Detection Laboratory

## DataMaster Evaluation Form

Issued: 7/10/2015  
Effective: 7/10/2015

Version: DMEF2015 R0  
Status: Active

Instrument # 100404

Date: 7/15/15

Analyst: Colleen O'Bryant

1. Reason for instrument evaluation

Instrument returned for preventative maintenance.

2. Visual Inspection

3. Troubleshooting/Repair

No repair required.

4. VOC (attached) 3.01 software

5. Internal Standard value  $\leq 3.00\%$  Yes  No

6. Verify software version

Return to Factory for Repair

Ready For Calibration

Ready For Certification

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100404

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/15/2015

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.70: 0.079  
LOT #: 07214080A1  
EXPIRATION: 04/01/2016  
TANK PRESSURE: 989 psi

BLANK TEST	0.000	15:09
INTERNAL STANDARD	VERIFIED	15:09
Xq = 0.088 (0.88%)		
EXTERNAL STANDARD	0.077	15:10
X[1] = 0.0772 (-0.0033) (-0.0017)		
BLANK TEST	0.000	15:10
EXTERNAL STANDARD	0.078	15:11
X[1] = 0.0775 (-0.0029) (-0.0016)		
BLANK TEST	0.000	15:12
EXTERNAL STANDARD	0.078	15:12
X[1] = 0.0775 (-0.0030) (-0.0015)		
BLANK TEST	0.000	15:13
EXTERNAL STANDARD	0.078	15:13
X[1] = 0.0776 (-0.0027) (-0.0015)		
BLANK TEST	0.000	15:14
EXTERNAL STANDARD	0.078	15:14
X[1] = 0.0782 (-0.0023) (-0.0010)		
BLANK TEST	0.000	15:15

Average = 0.0776  
Std Dev = 0.0004

## Diagnostic Check

### VERSIONS

DMT: 3.01  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.8°C	PASSED
(44.0°C - 52.0°C)	
Breath Tube = 46.9°C	PASSED
(38.0°C - 50.0°C)	

### SETTINGS

Lamp Voltage = 1.74 V	PASSED
(1.06 V - 2.48 V)	
Cooler Voltage = 1.65 V	PASSED
(1.01 V - 2.35 V)	
Bias Voltage = 82 V	PASSED
(49 V - 115 V)	
Chopper Freq = 522 Hz	PASSED
(475 Hz - 575 Hz)	
Barometer = 29.7 in	

### PUMP INFO

Flow Rate = 4.552 L/M	PASSED
(3.500 L/M - 6.500 L/M)	

### DETECTOR INFO

PUMP ON	PASSED
(0.004 V <= 0.010 V)	
PUMP OFF	PASSED
(0.003 V <= 0.010 V)	

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD PASSED

Xq = 0.088 0.98% (0.00% - 4.00%)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this \_\_\_\_\_ day of \_\_\_\_\_, 20 \_\_\_\_\_