

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100673

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 04/13/2025

External Standard Test Values

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 44.9°C	PASSED

PUMP INFO	
Flow Rate = 4.341 L/M	PASSED

DETECTOR INFO	
PUMP ON	PASSED
PUMP OFF	PASSED

RFI DETECTED

File only
CRF 5/12/25