

# Alaska Scientific Crime Detection Laboratory

## DataMaster Evaluation Form

Issued: 7/10/2015  
Effective: 7/10/2015

Version: DMEF2015 R0  
Status: Active

Instrument # 100691

Date: 8/12/15

Analyst: Colleen O'Bryant

1. Reason for instrument evaluation

Attempting to update software - screen is unresponsive

2. Visual Inspection

3. Troubleshooting/Repair

Screen is unresponsive + blacks following restarts and checks of all connections inside

4. VOC (attached) N/A

5. Internal Standard value  $\leq 3.00\%$  Yes  No  N/A

6. Verify software version  3.01

Return to Factory for Repair

Ready For Calibration

Ready For Certification