

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 6/11/2014  
Effective: 6/11/2014

Version: CIISF 2014 R0  
Status: Active

Date: 3/10/15 Instrument Serial Number: 100.381  
Instrument Location: ANCHORAGE TAIL  
Supervisor Name and Agency: S. DUNN - APD

Describe the Status Change or Issue with the Instrument:

- RESET OPTIONS / NON-DRINKING SUBJECT TEST / VOC

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 3-10-15 @ 1445

Out of Service Date/Time: N/A

For Use by SCDL

Additional Notes

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740

SUBJECT TEST

Alaska Department of Public Safety  
 DATAMASTER dmt: 100381

Date: 03/10/2015  
 Time: 14:37:38

SUBJECT NAME:  
 TEST TEST  
 LICENSE #: 00000

OPERATOR NAME:  
 STEVE DUNN  
 OPERATOR'S #: 4106

DEPT/AGENCY:  
 ANCI  
 CASE/REPORT: 15-000  
 TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION  
 NOMINAL: 0.080  
 TARGET AT 30.10: 0.080  
 LOT #: 11313080A1  
 EXPIRATION: 06/01/2015

| DIAGNOSTIC CHECK           | PASSED   | 14:38 |
|----------------------------|----------|-------|
| BLANK TEST                 | 0.000    | 14:38 |
| INTERNAL STANDARD          | VERIFIED | 14:38 |
| EXTERNAL STANDARD          | 0.080    | 14:39 |
| BLANK TEST                 | 0.000    | 14:39 |
| SUBJECT SAMPLE (Vol=2.66L) | 0.000    | 14:40 |
| BLANK TEST                 | 0.000    | 14:41 |
| EXTERNAL STANDARD          | 0.080    | 14:41 |
| BLANK TEST                 | 0.000    | 14:42 |

*S. DUNN*

