

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100690

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 04/19/2019

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.78: 0.080
LOT #: AG826101
EXPIRATION: 09/18/2020
TANK PRESSURE: 1131 psi

BLANK TEST	0.000	22:09
INTERNAL STANDARD	VERIFIED	22:09
EXTERNAL STANDARD	0.077	22:10
BLANK TEST	0.000	22:11
EXTERNAL STANDARD	0.077	22:11
BLANK TEST	0.000	22:12
EXTERNAL STANDARD	0.077	22:12
BLANK TEST	0.000	22:13
EXTERNAL STANDARD	0.076	22:13
BLANK TEST	0.000	22:14
EXTERNAL STANDARD	INTERFERENCE	22:14

Interference Detected

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 45.9°C	PASSED

PUMP INFO

Flow Rate = 4.307 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------