

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 10/3/2024

Instrument #: 100699

Instrument Location: JBER

Supervisor Name and Agency: Scott Gardner, JBER

Describe the Status Change or Issue with the Instrument:

Click or tap here to enter text.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA

Out of Service Date/Time: 10/03/2024

For Use by SCDL

Additional Notes

During in field instrument review of 07-19-2024 to 09-05-2024, it was determined that due to the number of Internal Standard Errors on Subject Tests, the instrument should be taken out of service for maintenance. Scott was notified via email. See attached. KS 10/03/2024

Email completed form to dps.scdl.toxicology@alaska.gov.

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

Swisher, Kaija L (DPS)

From: CANNON, BRIAN M CIV USAF PACAF 673 SFS/S3 <brian.cannon.4@us.af.mil>
Sent: Thursday, October 3, 2024 8:43 PM
To: Swisher, Kaija L (DPS)
Cc: GARDNER, SCOTT E CIV USAF PACAF 673 SFS/S3T
Subject: RE: 100699 DataMaster Removal from Service

You don't often get email from brian.cannon.4@us.af.mil. [Learn why this is important](#)

CAUTION: This email originated from outside the State of Alaska mail system. Do not click links or open attachments unless you recognize the sender and know the content is safe.

100699 is packaged and ready to be returned to the lab. I can bring it by Monday at about mid-day at the earliest. Is there a particular time that works best?



Brian M. Cannon

Police Officer Supervisor

JBER Police Department/673 SFS

8517 20th Street, 99506

From: GARDNER, SCOTT E CIV USAF PACAF 673 SFS/S3T <scott.gardner.14@us.af.mil>
Sent: Thursday, October 3, 2024 5:39 PM
To: CANNON, BRIAN M CIV USAF PACAF 673 SFS/S3 <brian.cannon.4@us.af.mil>
Cc: Swisher, Kaija L (DPS) <kaija.swisher@alaska.gov>
Subject: RE: 100699 DataMaster Removal from Service

Brian,

See Kaija's message below. Will need you to swap 100699 for another loaner or 100422/21 if it's ready. Can't remember which one is in the lab for repair. It was the one with the non-functioning touch screen. 100699 is a loaner and the pelican case is in the LETC storage closet. Appreciate you!

-Scott



Scott Gardner
Training Officer/Program Manager
JBER Police Department/673 SFS
8517 20th Street, JBER, AK 99506
Office: 907-552-3537
Cell: 907-350-6140
scott.gardner.14@us.af.mil

[S3 TRAINING SITE \(dps.mil\)](#)

From: Swisher, Kaija L (DPS) <kaija.swisher@alaska.gov>
Sent: Thursday, October 3, 2024 3:23 PM
To: GARDNER, SCOTT E CIV USAF PACAF 673 SFS/S3T <scott.gardner.14@us.af.mil>
Subject: [Non-DoD Source] 100699 DataMaster Removal from Service
Importance: High

You don't often get email from kaija.swisher@alaska.gov. [Learn why this is important](#)

Good afternoon,

During the infield instrument status review for the cycle of 07/19/2024-09-05-2024 it was determined that the instrument 100699 should be removed from service and replaced with a new instrument. This is due to the amount of internal standard errors on subject tests.

It is recommended that the instrument be swapped for a new one as soon as possible for maintenance. Please let me know when the instrument can be returned to the lab. As of today the instrument should not be used for subject tests.

I will fill out a change in status form for this instrument (100699) so no need to worry about that. Just be sure to fill one out for the replacement.

Please let me know if you have any questions.

Thank you!

Kaija Swisher
Forensic Scientist II – Chemistry Section
Alaska Scientific Crime Detection Laboratory
4805 Dr. MLK Jr. Ave., Anchorage, AK 99507
Work: 907-269-5686