

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 6/11/2014  
Effective: 6/11/2014

Version: CIISF 2014 R0  
Status: Active

Date: 9-1-17 Instrument Serial Number: 100350

Instrument Location: Jail - room # B105

Supervisor Name and Agency: P.E. Martin

Describe the Status Change or Issue with the Instrument:

- High rate of Interference Status messages
- Internal Standard Error

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: 9-1-17 @ 0921

**For Use by SCDL**

**Additional Notes**

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740