







## **In Field Instrument Review**

**10/24/2015 – 12/12/2015**

- 100342 – AST/Fairbanks – Instrument removed from service
  - (1) RFI, (1) Gas Flow Error, (2) Static Analysis Errors
- 100344 – AST/Girdwood - No action taken
  - RFI (5) 2/4 tests, Diagnostic Failed (1)
- 100346 – Fairbanks Airport Police – No action taken
  - RFI (1), Interference (2)
- 100349 – Ketchikan PD – Requested ILMO be switched
  - Standard Out of Range (4), Ambient Fail (1)
- 100351 – Anchorage PD – Out of service 11/18/2015 for detector overflow
  - Interference (1), Tank Interference (1), Blank Error (1), Internal Standard Error (1), Detector Overflow (1)
- 100352 – AST/Bethel – Out of service 11/24/2015
  - Blank Error (1)
- 100358 – UAA PD – No action taken
  - Gas Flow Error (2) 2/4 tests, external standard running ~0.004 low on VOC and subject tests, only 4 tests in last 60 days
- 100364 – Anchorage PD – No action taken
  - Interference (1), Tank Interference (9) – cluster between 12/09/2015 and 12/12/2015, Blank Error (1)
- 100367 – Kotzebue PD – Out of service 11/25/2015
  - Interference (2), Standard Out of Range (4)
- 100368 – AST/Willow – No action taken
  - Blank Error (2) – both on same test, not repeated
- 100370 – Homer PD – No action taken
  - RFI (3)
- 100376 – Metlakatla PD – No action taken
  - Blank Error (1), Ambient Fail (1)
- 100383 – Eielson AFB – In service 12/8/2015, out of service 12/14/2015
  - Pump Error (1), Blank Error (1)
- 100387 – Fairbanks PD – No action taken
  - Blank Error (1), Ambient Fail (1)
- 100391 – Wrangell PD – No action taken
  - RFI (1)
- 100395 – Anchorage PD – No action taken
  - Ambient Fail (1)
- 100398 – Seward PD – No action taken
  - RFI (6) 33% of tests, RFI followed by completed test
- 100404 – Anchorage PD – No action taken
  - Ambient Fail (1)
- 100421 – JBER – No action taken

- Blank Error (1), Standard Out of Range (1), Ambient Fail (2), Internal Standard Error (1) – all except blank error occurred with the same subject and haven't been duplicated.
- 100422 – JBER – No action taken
- RFI (1)
- 100423 – Fort Wainwright – No action taken
- RFI (1)
- 100424 – Eielson AFB – Out of service 12/8/2015
- RFI (5)
- 100665 – AST/St. Mary's – No action taken
- Diagnostic Failed (1) – followed by a completed test and not duplicated
- 100673 – NSBPD – No action taken
- Pump Error (1)
- 100677 – Bethel PD – No action taken
- (1) Static Analysis error – not replicated
- 100683 – AST/Glennallen – Monitoring instrument while in service
- RFI (2), Tank Interference (2) – on VOCs, 2 failed, one successful.
- 100688 – AST/Soldotna – Requested ILMO be switched
- RFI (1), Tank Interference (2), Standard Out of Range (2), Ambient Fail (1)
- 100690 – Sitka PD – No action taken
- Blank Error (1)
- 100695 – Wasilla PD – No action taken
- RFI (4), Pump Error (2)
- 100697 – Mat-Su West – Monitoring instrument while in service
- RFI (7) – 33%, Clustered in two weeks of November
- 100700 – Soldotna PD – No action taken
- RFI (1)