

Alaska Scientific Crime Detection Laboratory

DataMaster Repair Form

Effective: 3/20/2025

Version: 2.0

Instrument #: 100411

Analyst(s): Carlie Glaister

Start Date: 11/3/25

Reason for Repair

Internal standard errors

Troubleshooting/Repair

Cleaned internal standard & filters
with 150 proof alcohol.

There is a sheared off screw, this was left as
is due to 3 other screws being present to
close lid.

☒ VOC (attached)

☒ VOC Internal Standard value $\leq 3.00\%$

☒ Verify software version

☒ Non-Drinking Subject Test (attached)

☒ Ready for Adjustment

☐ Further Repair Needed

☐ Proceed to Close Out

☐ Ready for Calibration

☐ Return to Factory

End Date: 11/3/25

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100411

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 11/03/2025

External Standard Test Values

Diagnostic Check

VERSIONS

DMT: 3.02

PIC: 3.02

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 49.1°C

PASSED

(44.0°C - 52.0°C)

Breath Tube = 43.3°C

PASSED

(38.0°C - 50.0°C)

SETTINGS

Lamp Voltage = 1.70 V

PASSED

(1.02 V - 2.38 V)

Cooler Voltage = 1.82 V

PASSED

(1.09 V - 2.55 V)

Bias Voltage = 80 V

PASSED

(48 V - 112 V)

Chopper Freq = 511 Hz

PASSED

(475 Hz - 575 Hz)

Barometer = 29.4 in

PUMP INFO

Flow Rate = 4.002 L/M

PASSED

(3.500 L/M - 6.500 L/M)

DETECTOR INFO

PUMP ON

PASSED

(0.002 V <= 0.010 V)

PUMP OFF

PASSED

(0.002 V <= 0.010 V)

FILTER INFO

Filter 1

PASSED

Filter 2

PASSED

Filter 3

PASSED

INTERNAL STANDARD

FAILED

Xq = 0.133 38.26% (0.00% - 4.00%)

Diagnostic Check Failed