

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100681

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/27/2019

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.15: 0.078
LOT #: AG923401
EXPIRATION: 08/22/2021
TANK PRESSURE: 105 psi

BLANK TEST	0.000	17:10
INTERNAL STANDARD	VERIFIED	17:10
EXTERNAL STANDARD	0.078	17:10
BLANK TEST	0.000	17:11
EXTERNAL STANDARD	0.079	17:11
BLANK TEST	0.000	17:12

Tank Pressure Too Low

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 47.4°C	PASSED

PUMP INFO

Flow Rate = 4.314 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

PASSED