

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100670

Received From: Unalaska

Date: 8/6/24

Analyst: C. O'Bryant

Reason for instrument return:

Screen issues, modem issues + upcoming preventative maintenance

Change instrument and case location in JusticeTrax Breath Alcohol Database COB 2024.08.27

Ensure change in instrument status form is in JusticeTrax

Upload ^{COB 8/6/24} *unable to upload via modem. uploaded via USB - no files found.* ^{COB 8/6/24}

Update agency information and polling status in DMHost Calling Program

Check for VOCs

From Intoximeters

Intoximeters Repair Paperwork (attached)