

|   |   |                      |  |
|---|---|----------------------|--|
|  | State of Alaska Scientific Crime Detection Laboratory |                      |  |
|   | DataMaster Acceptance Form - DMAF2011R0               |                      |  |
|   | Issuing Authority - Nita Bolz, Scientific Director    |                      |  |
| Approval date: 03/01/11   | Effective date: 03/01/11                              | Archive date: Active |  |


DataMaster DMT Acceptance Form

Instrument # 130373

Date: 4/12/11

Analyst: Colleen O'Bryant

1. Check in instrument in to JusticeTrax Breath Alcohol Database
2. Visual Inspection
3. Update software
4. Print calibration factors (attached)
5. NPAS calibration paperwork (attached)
6. Barcode Instrument
7. State of Alaska Property ID# 10140659



## CALIBRATION FACTORS

-----  
Alaska Department of Public Safety  
DATAMASTER dnt: 100373  
-----

Date: 03/10/2011  
Time: 08:43:49

|     |   |          |        |    |             |
|-----|---|----------|--------|----|-------------|
| Ca  | = | 0.1000   |        |    |             |
| CAL | = | 0.979895 | 0.800  | <= | CAL < 1.200 |
| b1  | = | 0.0005   | 0.0000 | <= | b1 < 0.0040 |
| b2  | = | 0.0043   | 0.0010 | <= | b2 < 0.0100 |
| b3  | = | 0.0002   | 0.0000 | <= | b3 < 0.0040 |
| Xq  | = | 0.0821   | 0.0500 | <= | Xq < 0.2000 |
| a21 | = | 1.174340 | 1.050  | <= | a21 < 1.300 |
| a31 | = | 0.516911 | 0.300  | <= | a31 < 0.800 |



**National Patent  
Analytical Systems, Inc.**

*Certificate of Calibration*

100373

Instrument Serial Number

10180

Standard Reference Number

The above instrument has been calibrated and is hereby certified for use in determining alcohol breath concentration when used in accordance with the operating guidelines by a properly trained operator.

3-2-11

Date

Charles C. Townsend

Technician

Final page for report

\

VANCCLSAN01\JTRAX\CASEDOCS\B-100373\B-100373\_ECF-20140

7/3/2014