

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100379

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/29/2020

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.71: 0.079
LOT #: AG923401
EXPIRATION: 08/22/2021
TANK PRESSURE: 111 psi

BLANK TEST	0.000	01:29
INTERNAL STANDARD	VERIFIED	01:29
EXTERNAL STANDARD	0.080	01:29
BLANK TEST	0.000	01:30
EXTERNAL STANDARD	0.081	01:30
BLANK TEST	0.000	01:31
EXTERNAL STANDARD	0.079	01:31

Blank Error

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C	PASSED
Breath Tube = 40.5°C	PASSED

PUMP INFO

Flow Rate = 4.654 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

PASSED

File only DM 09/16/20