Alaska Scientific Crime Detection Laboratory Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 4/25/2024

Instrument #: 100669

Instrument Location: Palmer PD

Supervisor Name and Agency: Derek Walton - SCDL

Describe the Status Change or Issue with the Instrument:

The most recent In Field cycle showed 10 RFI status messages from 60 tests. I emailed the Breath Test Supervisor Luke Szipszky about looking into this. Nearly all breath tests resulting in RFI since 01/01/2024 have been followed by a completed test.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA

Out of Service Date/Time: NA

For Use by SCDL

Additional Notes

Click or tap here to enter text.

Email completed form to <u>dps.scdl.toxicology@alaska.gov</u>.

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

Walton, Derek J (DPS)

From:Lucas N. Szipszky <lszipszky@PalmerPolice.com>Sent:Thursday, April 25, 2024 3:41 PMTo:Walton, Derek J (DPS)Subject:Re: DataMaster RFIs

CAUTION: This email originated from outside the State of Alaska mail system. Do not click links or open attachments unless you recognize the sender and know the content is safe.

We had another Datamaster doing the same thing. I haven't been able to pinpoint a cause. I will trouble shoot some more.

Sent from my iPhone

On Apr 25, 2024, at 4:36 PM, Walton, Derek J (DPS) <derek.walton@alaska.gov> wrote:

CAUTION: This email originated from outside of the organization. Do not click links or open attachments unless you recognize the sender and know the content is safe.

Good afternoon,

During the most recent data review cycle, we noted that the Palmer PD DataMaster had 10 of 60 tests with RFI status messages. This is continuing above where we'd like it to be and I'm sure it's frustrating for the officers. Do you have any suggestions on what might improve this?

Derek J Walton

Chemistry Supervisor Alaska State Crime Lab 907-269-5628