

	State of Alaska Scientific Crime Detection Laboratory		
	DataMaster Calibration Form - DMCAF2011R1		
	Issuing Authority - Nita Bolz, Scientific Director		
	Approval date: 03/10/11	Effective date: 03/10/11	Archive date: Active

DataMaster DMT Calibration Form

Instrument# 100384Date: 4/26/11

1. Wet Bath Simulator Solution

- a. Level 0.100
 b. Manufacturer Bath
 c. Lot 10200
 d. Expiration Date 10/25/2012

2. Wet Bath Simulator

- a. Serial Number DR-4161
 b. Date Calibrated 8/25/10

3. Attach Calibration Factors

Breath Alcohol Analyst

C. O'Bryant

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100384

Date: 04/26/2011
Time: 08:39:23

SUPERVISOR NAME:
COLLEEN S O'BRYAN
CERTIFICATE #: 6237

LOT #: GUTH100
BOTTLE #: 10260
EXPIRATION: 10/25/2012

Ca	=	0.1000		
CAL	=	0.964532	0.800	<= CAL < 1.200
b1	=	0.0003	0.0000	<= b1 < 0.0040
b2	=	0.0054	0.0010	<= b2 < 0.0100
b3	=	0.0004	0.0000	<= b3 < 0.0040
Xg	=	0.0829	0.0500	<= Xg < 0.2000
a21	=	1.181030	1.050	<= a21 < 1.300
a31	=	0.525071	0.300	<= a31 < 0.800

003-2/2

Final page for report

\

VANCSCLSAN01\JTRAX\CASEDOCS\B-100384\B-100384_ECF-20140

7/3/2014