

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 16 Aug 2019 Instrument Serial Number: ~~100415~~ 100423 *JM 09/06/19*
Instrument Location: Fort Wainwright, AK
Supervisor Name and Agency: Rossignol, Corey FWMPD

Describe the Status Change or Issue with the Instrument:

Printer/memory issue with previous instrument, switched at request of
at recommendation of Crime Lab.
- DataMaster lost connection to printer. I reset the instrument, printer worked
but memory of all reports was blank. (all reports wiped from instrument memory)

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 16 Aug 19 0900

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740