

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 4/29/15 Instrument Serial Number: 100397

Instrument Location: ARD Jail

Supervisor Name and Agency: Colleen O'Bryant / SCDL

Describe the Status Change or Issue with the Instrument:

Instrument was placed in service by Chuck Fiske at ARD. After returning to the lab, Chuck was able to upload the instrument without issue. Four days later the instrument won't upload. I contacted Rick Steiding (ARD) and he checked the modem (line) connection as well as restarting the instrument. The line was connected to the box but the box was loose from the wall. I was able to upload the instrument but the location of the connection to the wall needs to be improved.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: N/A

Out of Service Date/Time: N/A

For Use by SCDL

Additional Notes

Nita Balz contacted Sgt Roy Dablano and requested the modem number connected to the DMT be relocated to a port on the same side of the room. CBS 7/29/15

The modem line has been relocated to the same side of the room as the DMT and functionally uploaded this morning. CBS 7/10/15

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740