

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100681

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 03/02/2021

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 27.93: 0.075  
LOT #: AG923401  
EXPIRATION: 08/22/2021  
TANK PRESSURE: 883 psi

BLANK TEST	0.000	09:09
INTERNAL STANDARD	VERIFIED	09:09
EXTERNAL STANDARD	0.072	09:10
BLANK TEST	0.000	09:10
EXTERNAL STANDARD	0.072	09:11
BLANK TEST	0.000	09:11
EXTERNAL STANDARD	0.072	09:12
BLANK TEST	0.000	09:13
EXTERNAL STANDARD	0.072	09:13
BLANK TEST	0.000	09:14
EXTERNAL STANDARD	0.072	09:14
BLANK TEST	0.000	09:15

Average = 0.0720  
Std Dev = 0.0000

## Diagnostic Check

VERSIONS  
DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.7°C PASSED  
Breath Tube = 44.2°C PASSED

### PUMP INFO

Flow Rate = 4.061 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED

### FILTER INFO


Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

### INTERNAL STANDARD

PASSED

I, Brandi M. Barnett, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.





Brandi M. Barnett  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this

9 day of 4, 20 21



Notary Public  
My Commission Expires With Office



DM 03/02/21