

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 11/10/2021

Version: 2.0

Instrument #: 100413

Received From: APD Training

Date: 06/23/22

Analyst: Derech Watton

Reason for instrument return:

Returned from training, unable to load ticket database,
fatal errors when selecting DMT icon

Not returned with case DMV 06/23/22

Change instrument ~~and case location~~ in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax NA

Upload NA

Update agency information and polling status in DMHost Calling Program NA

Check for VOCs NA

Reset options DMV 06/23/22
From Intoximeters

Intoximeters Repair Paperwork (attached)

Replaced SD card -
ticket database loads Ok
and no more fatal errors.
DMV 06/23/22