

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 1/13/2017
Effective: 1/17/2017

Version: DMCIF 2017 R0
Status: Active

Instrument S/N 100411

Received From North Pole PD

Date 03/18/20

Analyst Derek Walton

1. Reason for instrument return:

Filter wheel error, non-functional Cermetex box installed by agency supervisor. Instrument never went in service at North Pole PD

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Upload → known dead modem (Cermetex box) - NA DW 03/18/20

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs

From Intoximeters

7. Intoximeters Repair Paperwork (attached)