

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 7/10/2015
Effective: 7/10/2015

Version: DMCIF 2015 R1
Status: Active

Instrument S/N 100699

Received From AST Anchor Pt

Date 10/23/15

Analyst Brandi Barnett

1. Reason for instrument return:

Regulator issues (Clicking & Std out of Range Low)

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Upload

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs /A

From NPAS

7. NPAS Repair Paperwork (attached)