

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 10-11-15 Instrument Serial Number: 100383

Instrument Location: Anchor Point AST

Supervisor Name and Agency: Ronny Simmons / AST

Describe the Status Change or Issue with the Instrument:
taken out of service to replace w/ new DM.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 10/11/15 16:26 hrs COB 10/14/15

Out of Service Date/Time: 10-11-15 1400 hrs

For Use by SCDL Additional Notes
Attempted to remove from service for DM, new instrument had regulator issues. 100383 placed back into service. COB 10/14/15

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740