

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 1/20/16 Instrument Serial Number: 100671

Instrument Location: Fairbanks AST

Supervisor Name and Agency: Trooper Malik Jones / AST Fairbanks

Describe the Status Change or Issue with the Instrument:

In-Service

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 1/20/16 0910 hours

Out of Service Date/Time: _____

For Use by SCDL

Additional Notes

Email completed form to

For questions contact the Breath Alcohol Program at 907-269-5740