

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Issued: 7/10/2015
Effective: 7/10/2015

Version: DMEF2015 R0
Status: Active

Instrument # 100415

Date: 11/18/15

Analyst: CWF

1. Reason for instrument evaluation

Returned for preventative maintenance

2. Visual Inspection

3. Troubleshooting/Repair

None

4. VOC (attached)

5. Internal Standard value $\leq 3.00\%$ Yes No

6. Verify software version

Return to Factory for Repair

Ready For Calibration

Ready For Certification

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100415

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 11/18/2015

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.99: 0.080
LOT #: 09214080A2
EXPIRATION: 05/01/2016
TANK PRESSURE: 249 psi

BLANK TEST	0.000	14:50
INTERNAL STANDARD	VERIFIED	14:50
Xq = 0.097 (0.09%)		
EXTERNAL STANDARD	0.080	14:50
X[1] = 0.0797 (-0.0019) (-0.0006)		
BLANK TEST	0.000	14:51
EXTERNAL STANDARD	0.079	14:51
X[1] = 0.0786 (-0.0031) (-0.0006)		
BLANK TEST	0.000	14:52
EXTERNAL STANDARD	0.079	14:52
X[1] = 0.0790 (-0.0023) (0.0005)		
BLANK TEST	0.000	14:53
EXTERNAL STANDARD	0.079	14:53
X[1] = 0.0787 (-0.0024) (0.0002)		
BLANK TEST	0.000	14:54
EXTERNAL STANDARD	0.079	14:55
X[1] = 0.0789 (-0.0023) (0.0002)		
BLANK TEST	0.000	14:55

Average = 0.0792
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.01
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.1°C	PASSED
(44.0°C - 52.0°C)	
Breath Tube = 44.4°C	PASSED
(38.0°C - 50.0°C)	

SETTINGS

Lamp Voltage = 1.73 V	PASSED
(1.04 V - 2.42 V)	
Cooler Voltage = 1.73 V	PASSED
(1.03 V - 2.39 V)	
Bias Voltage = 79 V	PASSED
(47 V - 111 V)	
Chopper Freq = 526 Hz	PASSED
(475 Hz - 575 Hz)	
Barometer = 30.0 in	

PUMP INFO

Flow Rate = 4.852 L/M	PASSED
(3.500 L/M - 6.500 L/M)	

DETECTOR INFO

PUMP ON	PASSED
(0.003 V <= 0.010 V)	
PUMP OFF	PASSED
(0.001 V <= 0.010 V)	

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

Xq = 0.096 0.54% (0.00% - 4.00%)	PASSED
----------------------------------	--------

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this _____ day of _____, 20 _____