

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100381

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 03/18/2016

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 28.65: 0.077
LOT #: 01315080A2
EXPIRATION: 02/05/2017
TANK PRESSURE: 1095 psi

BLANK TEST	0.000	10:33
INTERNAL STANDARD	VERIFIED	10:33
EXTERNAL STANDARD	0.075	10:33
BLANK TEST	0.000	10:34
EXTERNAL STANDARD	0.074	10:35
BLANK TEST	0.000	10:35
EXTERNAL STANDARD	0.075	10:36
BLANK TEST	0.000	10:37
EXTERNAL STANDARD	0.074	10:37
BLANK TEST	0.000	10:38
EXTERNAL STANDARD	0.074	10:38
BLANK TEST	0.000	10:39

Average = 0.0744
Std Dev = 0.0005

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.0°C PASSED
Breath Tube = 48.0°C PASSED

PUMP INFO

Flow Rate = 4.565 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Charles R. Foster, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist III at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Acting Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.



Charles R. Foster
Acting Scientific Director
State Breath Alcohol Program

4/15/16

Subscribed and sworn before me this 15th day of April, 20 16



Nikki Roth, Notary Public
My Commission Expires With Office



003 4/14/16