

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100676

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 04/23/2018

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 27.34: 0.073  
LOT #: AG735001  
EXPIRATION: 12/16/2019  
TANK PRESSURE: 926 psi

BLANK TEST	0.000	14:11
INTERNAL STANDARD	VERIFIED	14:11
EXTERNAL STANDARD	0.084	14:11

Standard Out of Range

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.8°C	PASSED
Breath Tube = 48.1°C	PASSED

### PUMP INFO

Flow Rate = 3.781 L/M	PASSED
-----------------------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

### INTERNAL STANDARD

PASSED