

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 12/15/2017  
Effective: 12/15/2017

Version: CIISF 2017 R0  
Status: Active

Date: 05/03/2019 Instrument Serial Number: 100350

Instrument Location: 258d Fourth Ave. Kotzebue, AK 99752

Supervisor Name and Agency: Ofc. James Rea Kotzebue Police Department

Describe the Status Change or Issue with the Instrument:

Place Out of order due repeated error of "Filter 3 wont Zero"

---

---

---

---

---

---

---

---

---

---

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: 05/03/2019 @ 1800 

-----  
**For Use by SCDL**

**Additional Notes**

---

---

---

---

---

---

---

---

---

---

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740