

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 11/12/2022

Instrument #: 100692

Instrument Location: 6040 De Havilland Ave. Anchorage, AK 99502

Supervisor Name and Agency: Thomas Hoffman / Anchorage Airport Police and Fire

Describe the Status Change or Issue with the Instrument:

Fatal error occurred while operator was attempting use of the instrument, test was not completed or recorded. While troubleshooting, keyboard stopped working. Taken out of service until it could be corrected.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: [Click or tap here to enter text.](#)

Out of Service Date/Time: 11/12/2022

For Use by SCDL

Additional Notes

[Click or tap here to enter text.](#)

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor