

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 9/17/2019 Instrument Serial Number: 100690

Instrument Location: UNALASKA

Supervisor Name and Agency: Sgt. WOOD UNALASKA

Describe the Status Change or Issue with the Instrument:

TAKEN OUT OF SERVICE AS REQUESTED By the
CRIME LAB DUE TO STATUS ERROR MESSAGES.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 9/17/2019 1700 Hours

For Use by SCDL

Additional Notes

Internal standard errors DMV 09/26/19

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740