

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Issued: 7/10/2015
Effective: 7/10/2015

Version: DMEF2015 R0
Status: Active

Instrument # 100252

Date: 8/11/15

Analyst: Colleen O'Bryant

1. Reason for instrument evaluation

Possible 5-way valve issue. Attempting to
troubleshoot prior to calibrating.

2. Visual Inspection

3. Troubleshooting/Repair

Loaded new software and attempted to
perform the required VOC. DMT shut down due
to a fatal error, repeatedly. Unable to perform
any other actions

4. VOC (attached) N/A

5. Internal Standard value $\leq 3.00\%$ Yes No N/A

6. Verify software version 3.01

Return to Factory for Repair

Ready For Calibration

Ready For Certification