

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100352

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 09/26/2013

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 28.92: 0.077
LOT #: 02012080A1
EXPIRATION: 03/01/2014

BLANK TEST	0.000	10:29
INTERNAL STANDARD	VERIFIED	10:29
EXTERNAL STANDARD	0.077	10:30
BLANK TEST	0.000	10:31
EXTERNAL STANDARD	0.077	10:31
BLANK TEST	0.000	10:32
EXTERNAL STANDARD	0.077	10:32
BLANK TEST	0.000	10:33
EXTERNAL STANDARD	0.077	10:33
BLANK TEST	0.000	10:34
EXTERNAL STANDARD	0.077	10:34
BLANK TEST	0.000	10:35

Average = 0.0770
Std Dev = 0.0000

Diagnostic Check

VERSIONS

DMT: 2.00
PIC: 3.03
Modem: 2.3
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C PASSED
Breath Tube = 47.4°C PASSED

PUMP INFO

Flow Rate = 4.830 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:



- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

10/9/2013

Subscribed and sworn before me this 9th day of Oct, 20 13


Nikki Roth, Notary Public
My Commission Expires With Office



CRF
9/26/13