

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 6/14/15 Instrument Serial Number: 100349

Instrument Location: Mat-Su West

Supervisor Name and Agency: Pete Steen / MSW

Describe the Status Change or Issue with the Instrument:

Pete Steen emailed to let me know an officer had encountered a "fatal error" twice while trying to initiate a subject test. The officer said he was trying to select "DUI" from the test type drop down menu when the DMT gave a "fatal error" and would not allow him to run a test. Officer Steen restarted the instrument and was able to run a non-drinking subject test without issue.
CSB 6/15/15

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: N/A

Out of Service Date/Time: N/A

For Use by SCDL

Additional Notes

Unable to recreate issue in the field (test done by Trooper Pete Steen), instrument remains in service. CSB 6/15/15

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740