

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Issued: 7/10/2015
Effective: 7/10/2015

Version: DMEF2015 R0
Status: Active

Instrument # 100-413

Date: 8/11/15

Analyst: Colleen O'Bryen

1. Reason for instrument evaluation

Internal standard failure during VOC day as well as subject tests.

2. Visual Inspection N/A

3. Troubleshooting/Repair

Instrument restarted itself twice due to a fatal error, both were initiated by trying to access the records menu.

4. VOC (attached) N/A

5. Internal Standard value $\leq 3.00\%$ Yes No N/A

6. Verify software version 3.01

Return to Factory for Repair

Ready For Calibration

Ready For Certification