

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 8/18/2022

Instrument #: 100395

Instrument Location: Fairbanks International Airport Police

Supervisor Name and Agency: Vance Enderle – Fairbanks International Airport Police

Describe the Status Change or Issue with the Instrument:

This instrument was taken out of service due to an elevated internal standard value.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: Click or tap here to enter text.

Out of Service Date/Time: 8/18/2022 @ 0830 hrs

-----  
**For Use by SCDL**

Additional Notes

Click or tap here to enter text.

**Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)**  
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor