

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument  
State of Alaska

Serial #: 100673

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/23/2014

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 28.58: 0.076  
LOT #: 09214080A3  
EXPIRATION: 05/01/2016

BLANK TEST	0.000	09:39
INTERNAL STANDARD	VERIFIED	09:39
EXTERNAL STANDARD	0.073	09:39
BLANK TEST	0.000	09:40
EXTERNAL STANDARD	0.073	09:40
BLANK TEST	0.000	09:41
EXTERNAL STANDARD	0.073	09:41
BLANK TEST	0.000	09:42
EXTERNAL STANDARD	0.073	09:42
BLANK TEST	0.000	09:43
EXTERNAL STANDARD	0.074	09:44
BLANK TEST	0.000	09:44

Average = 0.0732  
Std Dev = 0.0004

## Diagnostic Check

### VERSIONS

DMT: 2.00  
PIC: 3.03  
Modem: 2.3  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 49.1°C PASSED  
Breath Tube = 44.9°C PASSED

### PUMP INFO

Flow Rate = 4.532 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED


### FILTER INFO

Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 5<sup>th</sup> day of Jan, 20 15

  
Nikki Roth, Notary Public  
My Commission Expires With Office



  
12/23/14