

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 9/14/2023

Instrument #: 100370

Instrument Location: University of Alaska Fairbanks

Supervisor Name and Agency: James W. Thomas , University of Alaska Fairbanks

Describe the Status Change or Issue with the Instrument:

Instrument 100370 was removed from service at 1455 hours to allow another DMT to be placed into service.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: [Click or tap here to enter text.](#)

Out of Service Date/Time: 09/14/2023 1455 hours

For Use by SCDL

Additional Notes

Blank errors and internal standard errors occurring. BMB 9/15/2023

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor